



CELL LAB Rat Anti-Human G-CSF

Cat. No.	Form	Quantity
731776	Purified (UNLB) Antibody	0.5 mg

For Research Use Only. Not for use in diagnostic procedures.

DESCRIPTION

Clone: BVD13-3A5
Immunogen: *E. coli*-expressed recombinant human G-CSF
Isotype: Rat IgG1 κ
Specificity: Human Granulocyte-Colony Stimulating Factor (G-CSF)^{1,2} (This is a neutralizing antibody).

APPLICATIONS

Enzyme-Linked Immunosorbent Assay (ELISA) Capture: BVD13-3A5 monoclonal antibody is useful as a capture antibody in a sandwich ELISA for quantifying human G-CSF protein levels.^{1,2} Purified BVD13-3A5 antibody should be paired with biotinylated BVD11-37G10 antibody (Cat. No. 731778) as the detection antibody, with purified recombinant human G-CSF as the standard.

Western Blotting: The purified BVD13-3A5 antibody is also useful for Western blotting applications. A concentration of 1-5 μ g/mL (in conjunction with AP-labeled Goat Anti-Rat Ig, Cat. No. 732647) has been found to enable detection of \leq 100 ng/lane of recombinant human G-CSF under reducing conditions.

CHARACTERIZATION

To ensure lot-to-lot consistency, each batch of product is tested to conform with characteristics of a standard reference reagent using a sandwich ELISA.

WORKING DILUTIONS

Other Applications: Since applications vary, determine the optimum working dilution of the product that is appropriate for your specific needs.

HANDLING AND STORAGE

- The purified (UNLB) antibody is supplied as 0.5 mg of purified immunoglobulin in 1.0 mL of 100 mM borate buffered saline, pH 8.2. No preservatives or amine-containing buffer salts added.
- Reagent is stable until the expiration date on the vial when stored at 2-8°C.

STATEMENT OF WARNINGS

1. Specimens, samples and all material coming in contact with them should be handled as if capable of transmitting infection and disposed of with proper precautions.
2. Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
3. Do not use reagent beyond the expiration date on the vial label.
4. Avoid microbial contamination of reagent or erroneous results may occur.
5. Use Good Laboratory Practice (GLP) when handling this reagent.

TRADEMARKS

The Beckman Coulter logo is a trademark of Beckman Coulter, Inc.

For additional information or if damaged product is received, contact your local Beckman Coulter Representative.

REFERENCES

1. Abrams J. "Immunoenzymetric assay of mouse and human cytokines using NIP-labeled anti-cytokine antibodies," published in *Current Protocols in Immunology*, eds., J Coligan, A Kruisbeek, D Margulies, E Shevach, and W Strober (New York: John Wiley and Sons. Unit 6.20, 1997).
2. Abrams JS, Roncarolo MG, Yssel H, Andersson U, Gleich GJ, and Silver JE. 1992. Strategies of anti-cytokine monoclonal antibody development: Immunoassay of IL-10 and IL-5 in clinical samples. *Immunol Rev*, 127:5-24.



Manufactured for:
Beckman Coulter, Inc.
4300 N. Harbor Blvd.
Fullerton, CA 92835
www.beckmancoulter.com

Printed in USA
Made in USA

© 2005 Beckman Coulter, Inc.
All Rights Reserved.

PN 733821-A