

IOTest® CD8β-ECD

REF 6607123 - 100 tests

PN 4238123-C



	CD8β-ECD
Specificity	CD8β
Clone	2ST8.5H7
Hybridoma	Myeloma NS1 x BALB/c spleen cells
Immunogen	Normal human T cells
Ig Chain	IgG2a
Species	Mouse
Source	Ascites fluid
Purification	Ion exchange or affinity chromatography
Fluorescence	Excites at 486-580 nm / Emits at 610-635 nm
Conjugation	ECD (Phycoerythrin-Texas Red®-X)
Molar Ratio	ECD/Protein 0.5-1.5
Scatter Detection	Forward and/or side

**For Research Use Only.
Not for use in diagnostic procedures.**

SPECIFICITY

2ST8.5H7 antibody recognizes an epitope of the CD8β molecule that depends upon the expression of both CD8α and CD8β chains.¹⁻³ CD8 is composed of two polypeptide chains α and β which form homodimeric (CD8 α/α) or heterodimeric (CD8 α/β) cell surface complexes.

Four distinct subpopulations of CD8⁺ cells have been identified based on the expression of CD8 α/α or CD8 α/β complexes:¹

- TCR αβ⁺ T cells, which are CD8 α⁺/β⁺
- TCR αβ⁺ T cells, which are CD8 α⁺/β⁻
- TCR γδ⁺ T cells, which are CD8 α⁺/β⁻
- NK cells, which are CD8 α⁺/β⁻

Moreover, both CD8 α/α and CD8 α/β complexes may coexist on the cell surface.

2ST8.5H7 immunoprecipitates the human complex CD8 α/β (30-32 kDa) from cell lines expressing the α and β cDNA.² 2ST8.5H7 antibody has been studied at the fifth International Workshop on Human Leukocyte Differentiation Antigens.^{3,4}

REAGENT

See table above.

REAGENT CONTENTS

Contact Beckman Coulter Customer Service to obtain the antibody concentration in the IOTest reagent.

The nonantibody reagents are 2 mg/mL bovine serum albumin in phosphate-buffered saline containing 0.1% sodium azide.

STATEMENT OF WARNINGS

1. This reagent contains 0.1% sodium azide. Sodium azide under acid conditions yields hydrazoic acid, an extremely toxic compound. Azide compounds should be flushed with running water while being discarded. These precautions are recommended to avoid deposits in metal piping in which explosive conditions can develop. If skin or eye contact occurs, wash excessively with water.
2. Specimens, samples and all material coming in contact with them should be handled as if capable of transmitting infection and disposed of with proper precautions.
3. Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
4. Do not use antibody beyond the expiration date on the label.

5. Do not expose reagents to strong light during storage or incubation.
6. Use Good Laboratory Practices (GLP) when handling reagent.
7. Harmful if swallowed.
8. After contact with skin, wash immediately with plenty of water.

STORAGE CONDITIONS AND STABILITY

This reagent is stable up to the expiration date when stored at 2-8°C. Do not freeze. Minimize exposure to light.

REAGENT PREPARATION

No reconstitution is necessary. This monoclonal antibody may be used directly from the vial. Bring reagent to 18-25°C prior to use.


PROCEDURE

This reagent is designed for flow cytometry. Assay volume: 10 μL per 5 x 10⁵ cells in one test, or per 100 μL whole blood. A wash is required to yield optimal results.

EXAMPLE DATA

The histogram shown is a monoparametric representation (Count versus Fluorescence Intensity) of lysed normal whole blood sample stained with CD8β-ECD monoclonal antibody (PN 6607123) and gated on lymphocytes.

Figure 1:
Acquisition with a COULTER® EPICS™ XL™/XL-MCL™ flow cytometer.



SELECTED RESEARCH REFERENCES

1. Terry, L.A., DiSanto, J.P., Small, T.N., Flomenberg, N., "Differential expression and regulation of the human CD8α and CD8β chains", 1990, Tissue Antigens, 35, 82-91.

2. DiSanto, J.P., Terry, L.A., Flomenberg, N., "Generation of anti-human CD8β-specific antibodies using transfectants expressing mixed-species CD8 heterodimers", 1991, J. Immunol. Methods, 141, 123-131.
3. Alcover, A., "CD8 cluster report", 1995, in Leucocyte Typing V, White Cell Differentiation Antigens, Schlossman, S.F., et al., Eds., Oxford Univ. Press, p. 353-354.
4. Knowles, R.W., "Immunochemical analysis of the T cell-specific antigens", 1985, in Leucocyte Typing II, 1, Human T lymphocytes, Reinherz, E.L., et al., Eds., Springer-Verlag New York inc., 1985, p. 260-288.

PRODUCT AVAILABILITY

IOTest CD8β-ECD Conjugated Antibodies
PN 6607123 - 100 tests - 10 μL/test

TRADEMARKS

Beckman Coulter Logo, COULTER, EPICS, IOTest, XL and XL-MCL are trademarks of Beckman Coulter, Inc.

Texas Red®-X is a trademark of Molecular Probes, Inc.

For additional information, or if damaged product is received, call Beckman Coulter Customer Service at 800-526-7694 (USA or Canada) or contact your local Beckman Coulter Representative.

 Beckman Coulter, Inc.
4300 N. Harbor Blvd.
Fullerton, CA 92835
www.beckmancoulter.com

Printed in USA
Made in USA

© 2007 Beckman Coulter, Inc.
All Rights Reserved.