

IOTest[®] Anti-TCR V β 14-FITC

PN IM1558 – 50 tests – 20 μ L / test – Clone CAS1.1.3

For Research Use Only. Not for use in diagnostic procedures.

SPECIFICITY

Human variable β 14 chain of the T-cell receptor, called TCRBV14S1 according to the nomenclature from Wei et al. (1) also referred to as TRBV27 (based on the IMGT gene nomenclature) (2, 3).

V β 14 is a single membered subfamily (PL8.1 and PH21 are identical) (4, 5).

The CAS1.1.3 monoclonal antibody stains from 2.2 to 5.6% of peripheral CD3 positive lymphocytes in normal blood.

The specificity of this antibody has been confirmed at the first Human TcR monoclonal Antibody Workshop in San Francisco in 1995 (6).

REAGENT

IOTest Anti-TCR V β 14-FITC Conjugated Antibody
PN IM1558 – 1 mL Liquid – 50 tests – 20 μ L / test.

Clone	CAS1.1.3
Isotype	IgG1, mouse
Immunogen	Mouse T-cell hybridoma transfected with human V β 14 gene segment.
Hybridoma Source	NS1 x Balb/c spleen cells
Purification	Ascites fluid ion exchange chromatography
Conjugation	FITC (Fluorescein isothiocyanate) is conjugated at 3 – 10 moles of FITC per mole of Ig.
Fluorescence	FITC (Green) Excites at 468 – 509 nm Emits at 504 – 541 nm
Buffer	2 mg/mL bovine serum albumin in phosphate-buffered saline containing 0.1% sodium azide.

APPLICATION

Studies of TCR V β 14 positive cells by flow cytometry.

STATEMENT OF WARNINGS

1. This reagent contains 0.1% sodium azide. Sodium azide under acid conditions yields hydrazoic acid, an extremely toxic compound. Azide compounds should be flushed with running water while being discarded. These precautions are recommended to

avoid deposits in metal piping in which explosive conditions can develop. If skin or eye contact occurs, wash excessively with water.

2. Specimens, samples and all material coming in contact with them should be handled as if capable of transmitting infection and disposed of with proper precautions.
3. Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
4. Do not use antibody beyond the expiration date on the label.
5. Do not expose reagents to strong light during storage or incubation.
6. Avoid microbial contamination of reagents or incorrect results might occur.
7. Use good laboratory practices when handling this reagent.

STORAGE CONDITIONS AND STABILITY

This reagent is stable up to the expiration date when stored at 2 – 8°C in the dark. Do not freeze.

EVIDENCE OF DETERIORATION

Any change in the physical appearance of this FITC-labeled reagent (clear, colorless to yellowish-green liquid) or any major variation in values obtained for control samples may indicate deterioration and the reagent should not be used.

REAGENT PREPARATION

No reconstitution is necessary. This monoclonal antibody may be used directly from the vial. Bring reagent to 18 – 25°C prior to use.

PROCEDURE

This reagent is designed for flow cytometry. A wash is required to yield optimal results. Assay volume: 20 μ L per 5 x 10⁵ cells in one test, or per 100 μ L whole blood. It is preferable to double stain the sample with another T-cell marker (CD3, CD4, CD8).

SELECTED RESEARCH REFERENCES

1. Wei, S., Charnley, P., Robinson, M.A., Concannon, P., "The extent of the human germline T-cell receptor V beta gene

segment repertoire", 1994, Immunogenetics, 40, 27-36.

2. Lefranc, M.P., Giudicelli, V., Ginestoux, C., Bodmer, J., Muller, W., Bontrop, R., Lemaitre, M., Malik, A., Barbie, V., Chaume D., "IMGT, the international ImMunoGeneTics database", 1999, Nucleic Acids Res., 27, 209-212.
3. Lefranc, M.P., "IMGT, the international ImMunoGeneTics database", 2003, Nucleic Acids Res., 31, 307-310.
4. Concannon, P., Pickering, L.A., Kung, P., Hood, L., "Diversity and structure of human T cell receptor beta chain variable region genes", 1986, Proc. Natl. Acad. Sci. USA, 83, 6598-6602.
5. Tillinghast, J.P., Behlke, M.A., Loh, D.Y., "Structure and diversity of the human T cell receptor beta chain variable region genes", 1986, Science, 233, 879-883.
6. Posnett, D.N., Romagné, F., Necker, A., Kotzin, B.L., Sekaly, R.-P., "First Human TcR Monoclonal Antibody Workshop", 1996, The Immunologist, 4, 5-8.

PRODUCT AVAILABILITY

IOTest Anti-TCR V β 14-FITC Conjugated Antibody
PN IM1558 – 1 mL Liquid – 50 tests – 20 μ L / test.

For additional information in the USA, call 800-526-7694.

Outside the USA, contact your local Beckman Coulter representative.

www.beckmancoulter.com

TRADEMARKS

Beckman Coulter, the Beckman Coulter logo, IOTest are registered trademarks of Beckman Coulter, Inc.

Manufactured by:
Immunotech SAS, a Beckman Coulter Company
130, avenue de Lattre de Tassigny, B.P. 177
13276 Marseille Cedex 9, France

©2007 Beckman Coulter, Inc.
All Rights Reserved

