

Monoclonal Antibody IOTest® CD14-PC5.5

PN A70204 - 50 tests - Liquid - 10 µL/test* - Clone RMO52

Analyte Specific Reagent.

Analytical and performance characteristics are not established.

SPECIFICITY

The CD14 antigen is a glycosyl-phosphatidylinositol-linked single-chain surface membrane glycoprotein with a molecular weight of 53-55 kDa.

CD14 is found on cells of myelomonocytic lineage. It is strongly expressed on monocytes, macrophages, and weakly on neutrophils (1, 2). It is also present on pleural phagocytic cells and on reticular dendritic cells, on Langerhans cells, and histiocytes (3, 4). CD14 is not expressed on B lymphocytes, T lymphocytes, NK cells, red blood cells and platelets.

The RMO52 monoclonal antibody does not react with T or B lymphocytes (1, 2). It has been assigned to the CD14 cluster of differentiation during the 6th International Workshop on Human Leucocyte Differentiation Antigens in Kobe, Japan, in 1996 (WS Code: MA62) (5).

REAGENT

IOTest CD14-PC5.5 Conjugated Antibody
PN A70204 - 50 tests - Liquid - 10 µL/test*

Clone	RMO52
Isotype	IgG2a, Mouse
Immunogen	Monocytes
Hybridoma	SP2/0 x Balb/c
Source	Ascites fluid
Purification	Ion exchange or affinity chromatography
Conjugation	R Phycoerythrin-Cyanin 5.5 (PC5.5)
Molar Ratio	PC5.5 / Ig : 0.5 - 1.5
Fluorescence	Excites at 488 nm Emits at 692 nm

REAGENT CONTENTS

This antibody is provided in phosphate-buffered saline, containing 0.1% sodium azide and 2 mg/mL bovine serum albumin.

STATEMENTS OF WARNING

1. This reagent contains 0.1% sodium azide. Sodium azide under acid conditions yields hydrazoic acid, an extremely toxic compound. Azide compounds should be flushed with running water while being discarded. These precautions are recommended to avoid deposits in metal piping in which explosive conditions can develop. If skin or eye contact occurs, wash excessively with water.

2. Specimens, samples and all material coming in contact with them should be handled as if capable of transmitting infection and disposed of with proper precautions.
3. Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
4. Do not use antibody beyond the expiration date on the label.
5. Do not expose reagents to strong light during storage or incubation.
6. Avoid microbial contamination of reagents or incorrect results might occur.
7. Use good laboratory practices when handling this reagent.

STORAGE CONDITIONS AND STABILITY

This reagent is stable up to the expiration date when stored at 2 – 8°C. Do not freeze.

REAGENT PREPARATION

No reconstitution is necessary. This monoclonal antibody may be used directly from the vial. Bring reagent to 18 – 25°C prior to use.

PRECAUTIONS

Due to the tandem structure of the fluorochrome, PC5.5 also emits light at 575 nm. This secondary emission peak varies from lot-to-lot of PC5.5. Therefore, for multi-color analysis, the compensation matrix should be carefully checked when changing the lot of a PC5.5-conjugate.

SELECTED RESEARCH REFERENCES

1. Todd III, R.F., Nadler, L.M., Schlossman, S.F., "Antigens on human monocytes identified by monoclonal antibodies", 1981, J. Immunol., 126, 1435-1442.
2. Todd, R.F., van Agthoven, A., Schlossmann, S.F., Terhorst, C., "Structural analysis of differentiation antigens Mo1 and Mo2 on human monocytes", 1982, Hybridoma, 1, 329-337.
3. Peters, J.H., Ruppert, J., Gieseler, R.K.H., Najjar, H.M., Xu, H., "Differentiation of human monocytes into CD14 negative accessory cells: do dendritic cells derive from the monocytic lineage?", 1991, Pathobiology, 59, 122-126.

4. Ziegler-Heitbrock, H.W.L., Ulevitch, R.J., "CD14: Cell surface receptor and differentiation markers", 1993, Immunol. Today, 14, 121-125.
5. Goyert, S.M., Cohen, L., Gangloff, S.C., Ashmun, R., Haeflner-Cavaillon, N., "CD14 Workshop panel report", 1997, Leucocyte Typing VI, White Cell Differentiation Antigens. Kishimoto, T., et al, Eds., Garland Publishing, Inc., 963-965.

TRADEMARKS AND PATENT

Beckman Coulter, the Beckman Coulter logo, COULTER, and IOTest are registered trademarks of Beckman Coulter, Inc.

Cy5.5 is a trademark of Amersham Biosciences, a GE Healthcare company. Cy5.5 is licensed under patent 5,268,486.

MANUFACTURED BY :

IMMUNOTECH SAS
a Beckman Coulter Company
130, avenue de Lattre de Tassigny
B.P. 177 - 13276 Marseille Cedex 9
France

For additional information in the USA, call 800-526-7694.

Outside the USA, contact your local Beckman Coulter representative.

www.beckmancoulter.com

Printed in France.
Made in France.

©2008 Beckman Coulter, Inc.

(*): 10 µL is the quantity of product sufficient to stain
5 x 10⁵ cells in a standard immunofluorescence assay