

Analyte Specific Reagent.

Analytical and performance characteristics are not established.

SPECIFICITY

The CD10 antigen is referred to as the Common Acute Lymphoblastic Leukemia Antigen (CALLA) (1, 2). It is a type II integral membrane protein of 100 kDa, identified as the human membrane-associated neutral endo-peptidase (EC3.4.24.11) (3, 4). It is expressed on uncommitted lymphoid precursors. CD10 expression is lost as cells enter the T lineage. In the B lineage, CD10 expression is lost later in ontogeny, as cells acquire surface Ig expression. It is also expressed on activated and proliferating B cells in the germinal center, and on neutrophils (6) as well as on bone marrow stromal cells. It is also expressed on a number of other cells of epithelial origin (5, 6).

The ALB1 mAb was studied during the first HLDA Workshop on Human Leucocyte Differentiation Antigens held in Paris, France, in 1984 (7).

REAGENT

IOTest CD10-PC5.5
Conjugated antibody
PN B16490 - 0.5 mL - Liquid - 10 µL/test

Clone	ALB1
Isotype	IgG1, Mouse
Immunogen	Human Leukemia cells
Hybridoma	NS1 x balb/c
Source	Ascites fluid or supernatant of in vitro cultured hybridoma cells.
Purification	Affinity chromatography
Conjugation	R Phycocerythrin-Cyanine 5.5 (PC5.5)
Molar Ratio	PC5.5 / Ig : 0.5 - 1.5
Fluorescence	Excites at 488 nm Emits at 692 nm

REAGENT CONTENTS

This antibody is provided in phosphate-buffered saline, containing 0.1% sodium azide and 2 mg/mL bovine serum albumin.

STATEMENTS OF WARNING

- This reagent contains 0.1% sodium azide. Sodium azide under acid conditions yields hydrazoic acid, an extremely toxic compound. Azide compounds should be flushed with running water while being discarded. These precautions are recommended to avoid deposits in metal piping in which explosive conditions can develop. If skin or eye contact occurs, wash excessively with water.
- Specimens, samples and all material coming in contact with them should be considered potentially infectious and disposed of with proper precautions.

- Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
- Do not use antibody beyond the expiration date on the label.
- Do not expose reagents to strong light during storage or incubation.
- Avoid microbial contamination of reagents or incorrect results might occur.
- Use good laboratory practices when handling this reagent.

STORAGE AND HANDLING CONDITIONS AND STABILITY

This reagent is stable up to the expiration date when stored at 2 – 8°C. Do not freeze. No reconstitution is necessary. This monoclonal antibody may be used directly from the vial. Bring reagent to 18 – 25°C prior to use.

PRECAUTIONS

Due to the tandem structure of the fluorochrome, PC5.5 also emits light at 575 nm. This secondary emission peak varies from lot-to-lot of PC5.5. Therefore, for multi-color analysis, the compensation matrix should be carefully checked when changing the lot of a PC5.5-conjugate.

SELECTED RESEARCH REFERENCES

- Shipp, M.A., Look, A.T., "Hematopoietic differentiation antigens that are membrane-associated enzymes: cutting is the key!", 1993, *Blood*, 82, 1052-1070.
- Lebien, T.W., McCormack, R.T., "The common acute lymphoblastic leukemia antigen (CD10) - Emancipation from a functional enigma", 1989, *Blood*, 73, 625-635.
- Letarte, M., Vera, S., Tran, R., Addis, J.B., Onizuka, R.J., Quackenbush, E.J., Jongeneel, C.V., McInnes, R.R., "Common acute lymphocytic leukemia antigen is identical to neutral endopeptidase", 1988, *J. Exp. Med.*, 168, 1247-1253.
- Shipp, M.A., Vijayaraghavan, J., Schmidt, E.V., Masteller, E.L., D'adamio, L., Hersh, L.B., Reinherz, E.L., "Common acute lymphoblastic leukemia antigen (CALLA) is active neutral endopeptidase 24.11 ("enkephalinase") : direct evidence by cDNA transfection analysis", 1989, *Proc. Natl. Acad. Sci. USA.*, 86, 297-301.
- Braun, Martin, P.J., Ledbetter, J.A., Hansen, J.A., "Granulocytes and cultured human fibroblasts express common acute lymphoblastic leukemia-associated antigens", 1983, *Blood*, 61, 718-725.

- Metzgar, R.S., Borowitz, M.J., Jones, N.H., Dowell, B.L., "Distribution of common acute lymphoblastic leukemia antigen in nonhematopoietic tissues", 1981, *J. Exp. Med.*, 154, 1249-1254.
- Boucheix, C., Perrot, J.Y., Mirshahi, M., Fournier, N., Billard, M., Giannoni, F., Bernadou, A., Rosenfeld, C., "Monoclonal antibodies against acute lymphoblastic leukemia differentiation antigens", 1984, *Leucocyte Typing I*, Bernard, A. et al., Springer Verlag, 671-672.

TRADEMARKS

Beckman Coulter logo and IOTest are trademarks of Beckman Coulter; Beckman Coulter logo, IOTest are registered in the USPTO and SIPO.

MANUFACTURED BY :

IMMUNOTECH SAS
a Beckman Coulter Company
130, avenue de Lattre de Tassigny
B.P. 177 - 13276 Marseille Cedex 9
France

For additional information in the USA, call 800-526-7694.

Outside the USA, contact your local Beckman Coulter representative.

www.beckmancoulter.com

Printed in France.
Made in France.

©2012 Beckman Coulter, Inc.