

IOTest® CD79a-PE

REF IM2221
100 tests; 2 mL
20 µL / test



IOTest
Conjugated Antibody

IVD



ENGLISH	Specifications
Specificity	CD79a
Clone	HM47
Hybridoma	NS1 x Balb/c
Immunogen	Synthetic peptide (amino acids 202 – 216) from the cytoplasmic part of the CD79a (Mb-1) protein
Immunoglobulin	IgG1, κ
Species	Mouse
Source	Ascites
Purification	Protein A affinity chromatography
Fluorochrome	Phycoerythrin (PE)
λ excitation	488 nm
Emission peak	575 nm
Buffer	PBS pH 7.2 plus 2 mg / mL BSA and 0.1% NaN ₃

USE

This fluorochrome-conjugated antibody permits the identification and numeration of cell populations expressing the CD79a antigen present in human biological samples by means of flow cytometry.

PRINCIPLE

This test is based on the ability of specific monoclonal antibodies to bind to the antigenic determinants expressed by leucocytes on the internal surface of the plasma membrane, in the cytoplasm or in the nucleus.

Following a membrane fixation/permeabilization step, internal staining of the leucocytes is performed by incubating the sample with the IOTest reagent. The leucocytes are then analyzed by flow cytometry.

The flow cytometer measures light diffusion and the fluorescence of cells. It makes possible the delimitation of the population of interest within the electronic window defined on a histogram, which correlates the orthogonal diffusion of light (Side Scatter or SS) and the diffusion of narrow-angle light (Forward Scatter or FS). Other histograms combining two of the different parameters available on the cytometer can be used as aids in the electronic gating stage depending on the application chosen by the user.

The fluorescence of the delimited cells is analyzed in order to distinguish the positively-stained events from the unstained ones. The results are expressed as a percentage of positive events in relation to all the events acquired by the electronic gating.

EXAMPLES OF CLINICAL APPLICATIONS

CD79a marker is used for the diagnostic of B-cell neoplasms as it permits to differentiate myeloid neoplasms from T-cell ones. In addition, CD79a is also used in the diagnosis of precursor B-acute lymphoblastic leukemia (pre-B-ALL) because many of these tumors are negative for other B-cell markers (1).

STORAGE AND STABILITY

The conjugated liquid forms must be kept at between 2 and 8°C and protected from light, before and after the vial has been opened.

Stability of closed vial: see expiry date on vial.

Stability of open vial: the reagent is stable for 90 days.

PRECAUTIONS

1. Do not use the reagent beyond the expiry date.
2. Do not freeze.
3. Let it come to room temperature (18 – 25°C) before use.
4. Minimize exposure to light.
5. Avoid microbial contamination of the reagents, or false results may occur.

6. Antibody solutions containing sodium azide (NaN₃) should be handled with care. Do not ingest and avoid contact with the skin, mucosa and eyes.

Moreover, in an acid medium, sodium azide can form the potentially dangerous hydrazoic acid. If it needs to be disposed of, it is recommended that the reagent be diluted in a large volume of water before pouring it into the drainage system so as to avoid the accumulation of sodium azide in metal pipes and to prevent the risk of explosion.

7. All blood samples must be considered as potentially infectious and must be handled with care (in particular: the wearing of protective gloves, gowns and goggles).
8. Never pipette by mouth and avoid all contact of the samples with the skin, mucosa and eyes.
9. Blood tubes and disposable material used for handling should be disposed of in ad hoc containers intended for incineration.

SAMPLES

Venous blood or bone marrow samples must be taken using sterile tubes containing an EDTA salt as the anticoagulant. The use of other anticoagulants is not recommended.

The samples should be kept at room temperature (18 – 25°C) and not shaken. The samples should be homogenized by gentle agitation prior to taking the test sample.

The samples must be analyzed within 24 hours of venipuncture.

METHODOLOGY

NECESSARY MATERIAL NOT SUPPLIED

- Sampling tubes and material necessary for sampling.
- Automatic pipettes with disposable tips for 20, 100 and 500 µL.
- Plastic haemolysis tubes.
- Calibration beads. For example: Flow-Set™ Fluorospheres (Ref. 6607007).
- To obtain optimal results, the following reagents are recommended:
 - IntraPrep™ Fixation/Permeabilization reagent (Ref. A07802 or A07803).
 - Leucocyte fixation reagent. For example: IOTest 3 Fixative Solution (Ref. A07800).
- Isotypic control: IOTest reagent. IgG1-PE (Ref. A07796).
- Buffer (PBS: 0.01 M sodium phosphate; 0.145 M sodium chloride; pH 7.2).
- Centrifuge.
- Automatic agitator (Vortex type).
- Flow cytometer.

PROCEDURE

For the following procedure, the use of the IntraPrep Fixation/Permeabilization reagent (Ref. A07802 or A07803) is recommended.

For each sample analyzed, in addition to the test tube, one control tube is required in which the cells are mixed with the isotypic control IgG1-PE (Ref. A07796).

For a blood sample, optimal staining is obtained using a number of leucocytes between 3 and 10 x 10³ cells / µL. If the leucocyte concentration is greater than 10 x 10³ cells / µL, dilute. In the case of a cell suspension, optimal staining is obtained with 5 x 10² cells / µL; dilute if necessary.

1. Add 50 µL of blood sampled into EDTA or 5 x 10⁵ cells to 2 tubes (one test tube, one control tube).
2. Add to each tube 100 µL of IntraPrep reagent 1 (Fixation).
3. Vortex the tubes vigorously for 3 to 5 seconds.
4. Incubate for 15 minutes at room temperature (18 – 25°C) protected from light.
5. Add 4 mL of PBS to each tube.
6. Centrifuge for 5 minutes at 300 x g at room temperature. Remove the supernatant by aspiration.
7. Add to each tube 100 µL of IntraPrep reagent 2 (Permeabilization). Let mix by diffusion. DO NOT VORTEX.
8. Incubate for 5 minutes at room temperature WITHOUT SHAKING.
9. Shake the tubes carefully and manually for 2 to 3 seconds.
10. Add 20 µL of the specific IOTest conjugated antibody to each test tube, and 20 µL of the appropriate isotypic control to each control tube. Vortex the tubes gently.
11. Incubate for 15 to 20 minutes at room temperature protected from light.
12. Add 4 mL of PBS and centrifuge at 300 x g for 5 minutes at room temperature.
13. Remove the supernatant by aspiration and resuspend the cell pellet in 0.5 to 1 mL of IOTest 3 Fixative Solution (Ref. A07800) at its working concentration (1X).
14. The preparations are ready for cytometric analysis.

NOTE: If the preparations are to be stored for more than 2 hours prior to cytometric analysis, it is advisable to store them at 2-8°C and protected from light. The preparations thus stored do not keep, however, for more than 24 hours.

PERFORMANCE

SPECIFICITY

The CD79a molecule is part of the CD79a / CD79b disulphide-linked heterodimer, non-covalently bound to surface immunoglobulins to form B cell receptors (BCR) (1). The expression of CD79a appears early in the ontogeny of B cells and its localization at the pro-B stage is therefore cytoplasmic. Later on, the CD79a forms part of the BCR. Its membrane expression persists up to the plasmocytic stage, the stage at which its localization once again becomes cytoplasmic (2).

MAb HM47 reacts with an intracytoplasmic epitope of the CD79a molecule (2). It was assigned to CD79a at the 5th HLDA Workshop on Human Leucocyte Differentiation Antigens held in Boston, USA, in 1993 (WS Code: cB017, Section B) (2).

LINEARITY

To test the linearity of staining of this reagent, a positive cell line (RAMOS) and a negative cell line (MOLT16) were mixed in different proportions with a constant final number of cells, so that the positive / negative cell line ratio of the mixture ranged from 0 to 100%.

Aliquots were stained using the procedure described above and linear regression between the expected values and the observed values was calculated.

Specificity	Linear regression	Linearity (R ²)
CD79a	Y = 1.0058 X - 1.3718	0.9991

EXPECTED VALUES

Each laboratory must compile a list of reference values based upon a group of healthy donors from the local population. This must be done by taking age, sex and ethnic group into account, as well as any other potential regional differences.

In our laboratories, the whole blood samples of 50 healthy adults were used. The results obtained for the count of the positive events of interest are given in the tables below:

Lymphocytes	Number	Mean (%)	SD	CV (%)
CD79a ⁺	50	11.92	5.26	44.2

INTRA-LABORATORY REPRODUCIBILITY

On the same day and using the same cytometer, 12 measurements of the positivity of a sample containing positive cells (peripheral blood from the same donor) were carried out. The results obtained are summarized in the following table:

Lymphocytes	Number	Mean (%)	SD	CV (%)
CD79a ⁺	12	16.82	0.49	2.89

INTER-LABORATORY REPRODUCIBILITY

On the same day and on the same sample containing positive cells (peripheral blood from the same donor), 12 measurements of the positivity were carried out by two technicians and the preparations analyzed using two different cytometers. The results obtained are summarized in the following tables:

Cytometer n° 1:

Lymphocytes	Number	Mean (%)	SD	CV (%)
CD79a ⁺	12	16.82	0.49	2.89

Cytometer n° 2:

Lymphocytes	Number	Mean (%)	SD	CV (%)
CD79a ⁺	12	17.28	0.51	2.93

LIMITATIONS OF THE TECHNIQUE

1. Flow cytometry may produce false results if the cytometer has not been aligned perfectly, if fluorescence spillover have not been correctly compensated for and if the regions have not been carefully positioned.
2. Accurate and reproducible results will be obtained as long as the procedures used are in accordance with the technical insert leaflet and compatible with good laboratory practices.
3. The conjugated antibody of this reagent is calibrated so as to offer the best specific signal/non-specific signal ratio. Therefore, it is important to adhere to the reagent volume/sample volume ratio in every test.
4. In the case of a hyperleucocytosis, dilute the specimen in PBS so as to obtain a value of approximately 5×10^9 leucocytes/L.
5. In certain disease states, such as severe renal failure or haemoglobinopathies, lysis of red cells may be slow, incomplete or even impossible. In this case, it is recommended to isolate mononucleated cells using a density gradient (Ficoll, for example) prior to staining.

MISCELLANEOUS

See the Appendix for examples and references.

TRADEMARKS

The BECKMAN COULTER logo, Beckman Coulter, Flow-Set, IntraPrep, IOTest are registered trademarks of Beckman Coulter, Inc.

MANUFACTURED BY:

IMMUNOTECH
a Beckman Coulter Company
130 avenue de Lattre de Tassigny
B.P. 177 – 13276 Marseille Cedex 9
France
Customer Services: (33) 4 91 17 27 27

www.beckmancoulter.com



APPENDIX TO REF IM2221

EXAMPLE

The graph below is a biparametric representation (Side Scatter versus Fluorescence Intensity) of a lysed normal whole blood sample. Staining is with IOtest CD79a-PE Conjugated Antibody (Ref. IM2221). All leucocytes are represented.

Acquisition and analysis were performed with a CYTOMICS FC 500 flow cytometer equipped with CXP Software

REFERENCES

1. Reth, M., Hombach, J., Wienands, J., Campbell, K.S., Chien, N., Justement, L.B., Cambier, J.C., "The B-cell antigen receptor complex", 1991, Immunol. Today, 6, 12, 196-200.
2. Engel, P., Wagner, N., Tedder, T.F., "CD79 workshop report", 1995, Leucocyte Typing V, White Cell Differentiation Antigens. Schlossman, S.F., et al., Eds., Oxford University Press, 667-670.

