

IOTest[®] Anti-TCR V β 11-PE

PN IM2290 – 50 tests – 20 μ L / test – Clone C21

For Research Use Only. Not for use in diagnostic procedures.

SPECIFICITY

Human variable β 11 chain of the T-cell receptor (TCR), also called TCRBV11S1 according to the nomenclature from Wei et al (1) and also referred to as TRBV25-1 (based on the IMGT gene nomenclature) (2, 3).

Two V β 11 sequences are described, PL3.12 (4) and PH15 (5). These sequences differ only in their leader sequence and therefore lead to the same mature protein.

The C21 recognizes the gene product of these sequences and stains 0.4% to 0.9% of peripheral CD3⁺ cells in normal blood.

An invariant V α 24/V β 11 T cell receptor is expressed in all individuals by clonally expanded CD4⁺ CD8⁻ T cells, reactive to bacterial antigens. This unique lymphocyte population restricted by the CD1d molecule recognition has been identified as the natural killer T (NKT) cells.

The specificity of this antibody has been confirmed at the First Human TcR Monoclonal Antibody Workshop in San Francisco in 1995 (6).

REAGENT

IOTest Anti-TCR V β 11-PE Conjugated Antibody
PN IM2290 – 1 mL Liquid – 50 tests – 20 μ L / test.

| | |
|---------------------|---|
| Clone | C21 |
| Isotype | IgG2a, mouse |
| Immunogen | Human T-cell hybridoma |
| Hybridoma | P3-X63-Ag.8.653 x SJL spleen cells |
| Source | Ascites fluid |
| Purification | ion exchange chromatography |
| Conjugation | R-phycoerythrin (PE) is conjugated at 0.5 – 1.5 moles of PE per mole of Ig. |

Excitation wavelength: 488 nm

Maximum emission wavelength: 575 nm

Main emission color: orange-red

Buffer 2 mg/mL bovine serum albumin in phosphate-buffered saline containing 0.1% sodium azide.

APPLICATION

Studies of TCR V β 11 positive cells by flow cytometry.

STATEMENT OF WARNINGS

1. This reagent contains 0.1% sodium azide. Sodium azide under acid conditions yields hydrazoic acid, an extremely toxic compound. Azide

compounds should be flushed with running water while being discarded. These precautions are recommended to avoid deposits in metal piping in which explosive conditions can develop. If skin or eye contact occurs, wash excessively with water.

2. Specimens, samples and all material coming in contact with them should be handled as if capable of transmitting infection and disposed of with proper precautions.
3. Never pipet by mouth and avoid contact of samples with skin and mucous membranes.
4. Do not use antibody beyond the expiration date on the label.
5. Do not expose reagents to strong light during storage or incubation.
6. Avoid microbial contamination of reagents or incorrect results might occur.
7. Use good laboratory practices when handling this reagent.

STORAGE CONDITIONS AND STABILITY

This reagent is stable up to the expiration date when stored at 2 – 8°C in the dark. Do not freeze.

EVIDENCE OF DETERIORATION

Any change in the physical appearance of this PE-labeled reagent (clear colorless to pinkish liquid) or any major variation in values obtained for control samples may indicate deterioration and the reagent should not be used.

REAGENT PREPARATION

No reconstitution is necessary. This monoclonal antibody may be used directly from the vial. Bring reagent to 18 – 25°C prior to use.

PROCEDURE

This reagent is designed for flow cytometry. A wash is required to yield optimal results. Assay volume: 20 μ L per 5 x 10⁵ cells in one test, or per 100 μ L whole blood. It is preferable to double stain the sample with another T-cell marker (CD3, CD4, CD8).

SELECTED RESEARCH REFERENCES

1. Wei, S., Charmley, P., Robinson, M.A., Concannon, P., "The extent of the human germline T-cell receptor V beta gene

segment repertoire", 1994, Immunogenetics, 40, 27-36.

2. Lefranc, M.P., Giudicelli, V., Ginestoux, C., Bodmer, J., Muller, W., Bontrou, R., Lemaitre, M., Malik, A., Barbie, V., Chaume D., "IMGT, the international ImMunoGeneTics database", 1999, Nucleic Acids Res., 27, 209-212.
3. Lefranc, M.P., "IMGT, the international ImMunoGeneTics database", 2003, Nucleic Acids Res., 31, 307-310.
4. Concannon, P., Pickering, L., Kung, P., Hood, L., "Diversity and structure of human T-cell receptor beta-chain variable region genes", 1986, Proc. Natl. Acad. Sci. USA, 83, 6598-6602.
5. Tillinghast, J.P., Behlke, M.A., Loh, D.Y., "Structure and diversity of the human T-cell receptor beta chain variable region genes", 1986, Science, 22, 879-883.
6. Posnett, D.N., Romagné, F., Necker, A., Kotzin, B.L., Sekaly, R.-P., "First Human TcR Monoclonal Antibody Workshop", 1996, The Immunologist, 4, 5-8.

PRODUCT AVAILABILITY

IOTest Anti-TCR V β 11-PE Conjugated Antibody
PN IM2290 – 1 mL Liquid – 50 tests – 20 μ L / test.

PE is licensed under patent 4,520,110

For additional information in the USA, call 800-526-7694.

Outside the USA, contact your local Beckman Coulter representative.

www.beckmancoulter.com

TRADEMARKS

Beckman Coulter, the Beckman Coulter logo, IOTest are registered trademarks of Beckman Coulter, Inc.

Manufactured by:
Immunotech SAS, a Beckman Coulter Company
130, avenue de Lattre de Tassigny, B.P. 177
13276 Marseille Cedex 9, France

©2007 Beckman Coulter, Inc.
All Rights Reserved

